


News from ID30B


Andrew McCarthy (EMBL)

Christoph Mueller-Dieckmann (ESRF)


EMBL



ID30B – Optical layout



ID30B – Experimental hutch






ID30B – Variable focusing capabilities

Energy range: 6-20 keV


Flux $\sim 5 \times 10^{12}$ phs/sec/mm² at 12.7 keV

Beam size: <40 μm² (apertures – 10, 20, 30, 50 and 75 μm²)

Vertical focusing, no horizontal focusing
(mirror unbent)



FlexHCD – A versatile sample changer



SPINE baskets (x12) = 120 samples





ESRF
SLS
BESSY
PETRAIII
ALBA


Unipucks (x11) = 176 samples




New sample holder types (36 samples)




8 PDB depositions (6 publications)




5LJK




5L6M




5LGA



5F33/5LQM



5JJ2



5HZA/5HZB

Workflows implemented in MxCuBE

The screenshot shows the MxCuBE software interface with a red box highlighting a list of implemented workflows. The interface includes a top menu bar (File, Instrumentation, Help), a toolbar with Collect, System, Feedback, Chat, and Logout buttons, and a left sidebar for Sample list, User, and Machine current information.

Sample centring section:

- Sample position: Omega: 310.00, Kappa: 0.00, Phi: 0.00
- Holder length: 23.067
- Sample video: Shows a bright-field image of a sample in a capillary, with a 100 μm scale bar. Buttons for Centre, Save, and Snapshot are visible.

Collection method section:

- Standard Collection
- Characterisation
- Helical Collection
- Energy Scan
- XRF Spectrum

Implemented Workflows (highlighted by a red box):

- X-ray Centring
- Mesh Scan
- Kappa Re-orientation
- Visual Re-orientation
- Helical characterisation
- Mesh and collect
- Enhanced characterisation
- Burn strategy
- Dehydration
- Trouble shooting

Bottom Left: A Firefox browser window titled "Passerelle EDM" showing the "Beamlne Expert System" interface with the MXPressE workflow diagram.

Bottom Right: ESRF and EMBL logos.

```
graph LR; Start((Start)) --> BeamlineSetup[Beamline setup]; BeamlineSetup --> PrepareMXPress[Prepare MXPress]; PrepareMXPress --> ExecuteMXPress[ExecuteMXPress]; ExecuteMXPress --> CommonErrorReporter[CommonErrorReporter]; CommonErrorReporter --> Director[Director]; Director --> Stop((Stop)); Director --> SetISPyBToSuccess[Set ISPyB to success]; SetISPyBToSuccess --> WorkflowEnded[Workflow ended with error messages]; WorkflowEnded --> SetISPyBWithErrors[Set ISPyB to success with error messages]; SetISPyBWithErrors --> Finished((Finished)); Finished --> Stop;
```

Fast mesh scans implemented in MxCuBE

mxcube (opid-30b)

File Instrumentation Help

Collect System Feedback Chat

User: opid-30b Group: Set Logout

Sample list

Mode: Manually mounted Show SC-details

Centring: Semi Automatic Synch ISPyB

manually-mounted

- X-ray Centring - 1
- Workflow task - 0

Diffraction (Imaging) (macro-molecules)

Use fast mesh (beta) true

Exposure time 0.02

Total oscillation range 1.0

Transmission 100.0

Continue

Stop Pause

Machine current 155.5 mA

7/8 multibunch 00:16

Flux +0.00 ph/s

Energy Current: 13.5651 keV
Move to: keV

Resolution Current: 2.999 Å
Move to: Å

Transmission Current: 100.00%
Set to: Filters

Beamstop distance 38.0

Cryo 99.99 K

Safety shutter closed

Fast shutter closed

Beamstop out

Capillary out

Current users


- Selecting gives control
- Allow timeout control
- Take control

My name: bacon

[2017-02-05 08:43:39] Workflow: phiz: 0.258
[2017-02-05 08:43:39] Workflow: phiy: -1.164
[2017-02-05 08:43:40] Workflow: waiting for input, verify parameters and press continue.

The screenshot shows the MxCuBE software interface. On the left, there's a 'Sample list' panel with 'Manually mounted' selected. Below it, under 'X-ray Centring', 'X-ray Centring - 1' is checked. In the bottom left, there's a 'Diffraction (Imaging) (macro-molecules)' panel with several parameters: 'Use fast mesh (beta)' is checked and circled in red; 'Exposure time' is set to 0.02; 'Total oscillation range' is 1.0; and 'Transmission' is 100.0. At the bottom left are 'Stop' and 'Pause' buttons. On the right, there's a large panel for 'Machine current' showing 155.5 mA, and other controls for 'Energy', 'Resolution', 'Transmission', 'Beamstop distance', 'Cryo', 'Safety shutter', 'Fast shutter', 'Beamstop', and 'Capillary'. At the bottom right, there's a 'Current users' section with checkboxes for 'Selecting gives control', 'Allow timeout control', and 'Take control', with 'My name: bacon' entered. The bottom of the screen shows some log messages from the workflow.

Fast mesh scans implemented in MxCuBE



EMBL



Fast mesh scans implemented in MxCuBE

mxcube (opid-30b)

File Instrumentation Help

Collect System Feedback Chat

User: opid-30b Group: Set Logout

Sample list

Mode: Manually mounted Show SC-details

Centring: Semi Automatic Synch ISPyB

Sample centring Characterisation

Characterisation Results

Indexing summary: Selected spacegroup: P222

a (Å)	b (Å)	c (Å)	alpha (°)	beta (°)	gamma (°)
60.687	63.963	70.141	90.000	90.000	90.000

Collection plan strategy: resolution limit is set by the radiation damage

Wedge	Subwedge	Start (°)	Width (°)	No Images	Exp time (s)	Max res (Å)	Rel trans (%)	Distance (mm)
1	1	57.00	0.05	2280	0.020	2.00	2.83	433.19

Minimal oscillation ranges for different completenesses

Wilson plot B-factor = 21.8 Å²

Relative Intensity total vs. Dose, DI/2 ≈ 7.63 MGy

Maximal oscillation width to avoid overlapped spots

Diffraction Plan

Forced space group	Anomalous data	Aimed multiplicity	Aimed completeness	Aimed I/sigma at highest res.	Aimed resolution (Å)	Min osc. width
	False	Default (optimized)	0.99	2.00	0.00	Default

Image quality indicators

File	Dozor score (1)	Tot integr signal (2)	Spot total	In-Res Total	Good Bragg	Ice Rings	Meth 1 Res	Meth 2 Res	Max unit cell
ref-Trypsin_1_0001.cbf	10.1	1611	24	24	22	0	3.37	NA	92.4
ref-Trypsin_1_0002.cbf	126.5	236894	294	274	232	1	2.37	2.09	104.6
ref-Trypsin_1_0003.cbf	75.0	132342	355	322	312	1	2.08	2.07	104.2
ref-Trypsin_1_0004.cbf	141.3	329322	312	306	270	0	2.55	2.07	104.6

1. Dozor score: criteria of diffraction signal strength that uses intensities over background vs resolution. Popov 2014, to be published.
 2. Total integrated signal, spot total etc: results from cctbx Spotfinder

Collect Queue Pause View parameters

Machine current 202.1 mA
 Flux: 6.36e+11 ph/s
 Energy 13.6561 keV
 Current: 0.914 Å
 Move to: keV Å
 Resolution 2.002 Å
 Current: 420.44 mm
 Move to: Å
 Transmission 100.00%
 Set to: Filters
 Beamstop distance 21.0
 Cryo 99.99 K
 Safety shutter opened
 Fast shutter closed
 Beamstop in
 Capillary in
 Current users
 Selecting gives control
 Allow timeout control
 Take control
 My name: bacon

(2017-02-05 09:03:24) Characterisation: Strategy: total no images 2280, total exposure time 45.6 [s]
 (2017-02-05 09:03:24) Characterisation: Strategy calculation successful.
 (2017-02-05 09:03:31) Characterisation completed.




EMBL



ID30B – MD2S allows plate screening capability

Data collection




ID30B – SBS plate format supported


Crystal Direct™



CrystalQuick™ X



In Situ-1™



Other SBS plate formats can be accommodated but need templates from users

In situ plate data collection in MxCuBE

mxcube (mx-1743)

File Instrumentation Help

Collect System Feedback Chat

User: mx-1743 Group: Set Logout

Sample list

Mode: Plate Show SC-details

Centring: No Centring Synch ISPyB

Sample centring

Sample position

Omega: 316.36 Kappa: 0.00 Phi: 0.00

Holder length: 32.500

Sample video

Back Light: 0.60 Focus: -0.729 Front Light: 0.0 Zoom: 5

Centre Save Snapshot

Centre beam Quick realign

100 µm 100 µm

Point no. 2 selected Aperture diameter: 50

Collection method

Standard Collection

Acquisition

Oscillation range: 0.1 First image: 1

Oscillation start: 316.36 Number of images: 10

Kappa: 0.0 Phi: 0.0

Detector mode:

Exposure time (s): 0.037

Energy (keV): 12.7 MAD

Resolution Å: 1.997

Transmission (%): 100.0

Inverse beam Subwedge size:

Shutterless

Data location

Folder: /data/visitor/mx1743/id30b/20151104/RAW_DATA

/XyA/A3-2

File name: xyla_23_####.cbf

Browse Prefix xyla

Run number 23

Processing

No. residues: 200 Space group:

Unit cell:
a: 0 b: 0 c: 0
α: 0 β: 0 γ: 0

Safety shutter closed

Fast shutter closed

Beamstop in

Capillary unknown

Current users

Selecting gives control

Allow timeout control

Ask for control My name: bacon

Add to queue

2015-11-04 11:49:00 Asking for input files writing

2015-11-04 11:49:00 Preparing acquisition, start=314.730000, wedge size=10

2015-11-04 11:49:04 Collection completed

Machine current 185.4 mA uniform multibunch 08:51

Flux: +0.00 ph/s

Energy

Current: 12.7000 keV 0.976 Å

Move to: keV Å

Resolution

Current: 1.997 Å 391.52 mm

Move to: Å


Transmission

Current: 100.00% Set to: Filters

ESRF EMBL

MX-1743 (EMBL)

- CrystalDirect plates from the HTX lab
 - Test of diffraction quality at RT
 - Highest resolution is 2.5 Å (cryo)
 - Best we saw was ~3 Å (maybe 2.8 Å)



Thaumatin crystals (12.7 keV)

Extended ISPyB

exi.esrf.fr/mx/#/mx/datacollection/session/54151/main

Search

Library:Main - Intranet Exit, removing your data...

ExiMX Extended ISPyB for MX BETA

Version: 0.9.6 Released: 2016/11/15

Home Shipment Proteins and Crystals Prepare Experiment Data Explorer Offline Data Analysis Help

search by protein acronym

New Tab

Run #4 osc Sep 9, 2016 5:16:44 PM /data/id30b/inhouse/opid30b/20160909/Raw_Data/Thaumatin/D7_1

Workflow

Type	OSC	P 4 21 2	Completeness	Res.	Rmerge	
Res. (corner)	1.5 Å (1.18 Å)	Inner	91%	4.1	0.1	
Sample		Outer	100%	1.5	0.7	
Prefix	collect	Overall	98%	1.5	0.1	
Images	500	Phi range	0.1 °	cell A	cell B	cell C
Transmission	9.82789	Phi start (total)	340° (50°)	58.5689	58.5689	151.592
Flux start	3.5e+10 ph/sec	Exposure Time	0.02 s	Alpha	Beta	Gamma
		Flux end	3.39e+10 ph/sec	90	90	90

Summary Beamline Parameters Data Collections 1 Sample Results 8 Workflow

Run #3 osc Sep 9, 2016 5:12:05 PM /data/id30b/inhouse/opid30b/20160909/Raw_Data/Thaumatin/D7_1

Workflow

Type	OSC	P 4 21 2	Completeness	Res.	Rmerge	
Res. (corner)	1.5 Å (1.18 Å)	Inner	93%	4.2	0.1	
Sample		Outer	99%	1.6	0.8	
Prefix	collect	Overall	99%	1.6	0.1	
Images	500	Phi range	0.1 °	cell A	cell B	cell C
Transmission	9.82789	Phi start (total)	340° (50°)	58.5727	58.5727	151.579
Flux start	3.51e+10 ph/sec	Exposure Time	0.02 s	Alpha	Beta	Gamma
		Flux end	3.52e+10 ph/sec	90	90	90

Summary Beamline Parameters Data Collections 1 Sample Results 15 Workflow

Workflow

Type	OSC	P 4 21 2	Completeness	Res.	Rmerge	
Res. (corner)	1.5 Å (1.18 Å)	Inner	93%	4.2	0.1	
Sample		Outer	99%	1.6	0.8	
Prefix	collect	Overall	99%	1.6	0.1	
Images	500	Phi range	0.1 °	cell A	cell B	cell C
Transmission	9.82789	Phi start (total)	340° (50°)	58.5727	58.5727	151.579
Flux start	3.51e+10 ph/sec	Exposure Time	0.02 s	Alpha	Beta	Gamma
		Flux end	3.52e+10 ph/sec	90	90	90

Summary Beamline Parameters Data Collections 1 Sample Results 15 Workflow

Thaumatin crystals


Data collection statistics

Wavelength (Å)	0.9763	0.7085
Phs/sec (Single bunch 4x10 mA and 20 μm aperture)	3.5×10^{10}	9.3×10^{10}
Exposure time (s)	0.02	0.02
Oscillation range (degrees)	50	50
Total dose (MGy) – flux2dose (Sasha)	0.42	0.53
Space group	P4 ₁ 2 ₁ 2	P4 ₁ 2 ₁ 2
Cell Dimensions a, b, c (Å)	58.6, 58.6, 151.6, 90, 90, 90	58.6, 58.6, 151.5, 90, 90, 90
Resolution (Å) (final shell)	50-1.5 (1.53-1.5)	50-1.5 (1.53-1.5)
Observed Reflections	149,639 (7,107)	153,095 (7,675)
Unique Reflections	42,124 (2,031)	40,896 (2,027)
Completeness (%) (final shell)	98.2 (98.4)	95.1 (98.2)
R_{meas} (%) (final shell)	14.2 (137)	12.6 (111)
<I/σ(I)> (final shell)	5.7 (1.0)	7.1 (1.3)

Model quality indicators

R_{cryst} (%) / R_{free} (%)	16.5/18.6	16.8/18.8
rms deviations, bonds (Å)/angles (°)	0.008/0.97	0.008/0.97

Thaumatin (12.7 keV) – tartaric acid



Fo-Fc omit map contoured at 5σ

Thaumatin crystals – S-SAD

Extended ISPyB + | i | exi.esrf.fr/mx/#/mx/datacollection/session/54178/main | C | Search | ☆ | ☰

Last Visited 1 Library:Main - Intranet 2 Exit, removing your da...

ExiMX Extended ISPyB for MX BETA Version: 0.9.6 Released: 2016/11/15

Home Shipment ▼ Proteins and Crystals ▼ Prepare Experiment Data Explorer ▼ Offline Data Analysis ▼ Help ▼ search by protein acronym



New Tab Log out OPID30b

Run #3 OSC Sep 12, 2016 10:12:13 AM
/data/id30b/inhouse/opid30b/20160912/Raw_DATA/Thaumatin7keV/D6_2x2

Workflow Protein Sample Prefix Images Transmission Flux start

Type	OSC	P 41 21 2	Completeness	Res.	Rmerge
Res. (corner)	2 Å (1.71 Å)	Inner	79%	7.7	6.9
		Outer	15	1.7	42.5
		Overall	77%	1.7	7.2
Wavelength	1.772 Å	cell A	58.5	58.5	151.6
Phi range	0.1 °	cell B	Alpha	Beta	Gamma
Images	500	cell C	90	90	90
Transmission	31.211				
Exposure Time	0.02 s				
Flux start	1.09e+11 ph/sec				
Flux end	1.1e+11 ph/sec				

Summary Beamline Parameters Data Collections 1 Sample Results 14 Work




  

Run #2 OSC Sep 12, 2016 10:11:09 AM
/data/id30b/inhouse/opid30b/20160912/Raw_DATA/Thaumatin7keV/D6_2x2

Workflow Protein Sample Prefix Images Transmission Flux start

Type	OSC	P 41 21 2	Completeness	Res.	Rmerge
Res. (corner)	2 Å (1.71 Å)	Inner	80%	8.1	5.7
		Outer	52%	1.8	28.4
		Overall	86%	1.8	6.4
Wavelength	1.772 Å	cell A	58.5	58.5	151.6
Phi range	0.1 °	cell B	Alpha	Beta	Gamma
Images	500	cell C	90	90	90
Transmission	31.211				
Exposure Time	0.02 s				
Flux start	1.12e+11 ph/sec				
Flux end	1.12e+11 ph/sec				

Summary Beamline Parameters Data Collections 1 Sample Results 12 Work

Thaumatin crystals – S-SAD

XSCALE merge of 4 data sets

Resolution (Å)	Completeness (%)	I/SigI	R-meas (%)	CC(1/2)	Anomal Corr	SigAno	Nano
15	97.50	25.71	9.90	98.7*	32	1.454	16
8	100.00	25.82	9.40	99.2*	37*	1.581	158
6	99.50	23.24	9.50	99.3*	45*	1.633	285
4	100.00	25.18	7.90	99.3*	32*	1.177	1271
3.5	99.80	23.38	7.20	99.4*	19*	0.974	919
3	100.00	21.58	7.20	99.6*	21*	0.965	1675
2.5	100.00	16.58	8.90	99.4*	15*	0.899	3341
2.2	100.00	12.52	12.00	99.1*	11*	0.896	3754
2	100.00	8.91	17.10	98.3*	5	0.875	3939
1.9	87.60	5.85	24.10	95.5*	6	0.908	2111
total	98.20	14.46	8.90	99.6*	14*	0.943	17469

Thaumatin crystals – S-SAD

Data collection statistics (merge 4 partial datasets)

Wavelength (Å)	1.7712
Phs/sec (Single bunch 4x10 mA and 20 μm aperture)	1.1 x 10 ¹¹
Exposure time (s)	0.02
Oscillation range (degrees)	50
Total dose (MGy) – <i>flux2dose (Sasha)</i>	0.42
Space group	P4 ₁ 2 ₁ 2
Cell Dimensions a, b, c (Å)	58.6, 58.6, 151.6, 90, 90, 90
Resolution (Å) (final shell)	50-1.9 (1.95-1.9)
Observed Reflections	220,045 (8,883)
Unique Reflections	21,397 (1,349)
Anomalous Completeness (%) (final shell)	99.2 (88.5)
R _{meas} (%) (final shell)	9.3 (27.9)
<I/σ(I)> (final shell)	18.1 (6.3)

Model quality indicators





R _{cryst} (%) / R _{free} (%)	16.5/18.6
rms deviations, bonds (Å)/angles (°)	0.008/0.97

Thaumatin (7 keV) – Anomalous map




S anomalous difference peaks using experimental anomalous differences and model phases contoured at 5σ

Lysozyme-Benzamidine crystals

Save Comments		View All Summary info		View All DataCollection					
Exp. Type	Image Prefix	Run #	Parameters	Results	Image Thumbnail	Crystal snapshot	Graph	Comments	
<p>Start time: 15:22:56 11-07-2016 (1 Item)</p> <p>OSC lysozyme_benz 2 Nb images: 250 Exp. time: 0.037 s Phi range: 0.10 ° Flux: 2.17E10 ph/sec Detector resolution: 1.50 Å Transmission: 1.00 Wavelength: 0.912 Å Total expo time: 9.25 s</p> <p>EDNA_proc grenades_fastproc grenades_parallelproc Space Group: P 4 2 2 Completeness:</p>  									
<p>Start time: 15:21:34 11-07-2016 (1 Item)</p> <p>OSC lysozyme_benz 1 Nb images: 250 Exp. time: 0.037 s Phi range: 0.10 ° Flux: 2.17E10 ph/sec Detector resolution: 1.50 Å Transmission: 1.00 Wavelength: 0.912 Å Total expo time: 9.25 s</p> <p>EDNA_proc grenades_fastproc grenades_parallelproc Space Group: P 4 2 2 Completeness:</p>  									
<p>Start time: 15:19:24 11-07-2016 (1 Item)</p> <p>OSC lysozyme_benz 5 Nb images: 250 Exp. time: 0.037 s Phi range: 0.10 ° Flux: 2.17E10 ph/sec Detector resolution: 1.50 Å Transmission: 1.00 Wavelength: 0.912 Å Total expo time: 9.25 s</p> <p>EDNA_proc grenades_fastproc grenades_parallelproc Space Group: P 41 21 2 Completeness:</p>  									
<p>Start time: 15:18:31 11-07-2016 (1 Item)</p> <p>OSC lysozyme_benz 3 Nb images: 250 Exp. time: 0.037 s Phi range: 0.10 ° Flux: 2.2E10 ph/sec Detector resolution: 1.50 Å Transmission: 1.00 Wavelength: 0.912 Å Total expo time: 9.25 s</p> <p>EDNA_proc grenades_fastproc grenades_parallelproc Space Group: P 41 21 2 Completeness:</p>  									

Lysozyme-Benzamidine crystals

- 9 datasets collected: (24° or 36°; 1% transmission; 0.037s; 0.1° per image)
- 8 merged using (HCA)




Data collection statistics

Wavelength (Å)	0.91165
Space group	P4 ₃ 2 ₁ 2
Cell Dimensions a, b, c (Å)	79.4, 79.4, 38.0, 90, 90, 90
Molecules per asymmetric unit	1
Resolution (Å) (final shell)	50-1.5 (1.53-1.5)
Observed Reflections	324,244 (5,095)
Unique Reflections	18,246 (847)
Completeness (%) (final shell)	91.7 (87.5)
R_{meas} (%) (final shell)	8.6 (48.6)
<I>/<math>\sigma(I)</math> (final shell)	21 (3.4)

Model quality indicators

R_{cryst} (%) / R_{free} (%)	16.5/17.8
rms deviations, bonds (Å)/angles (°)	0.01/1.1



Lysozyme-Benzamidine crystals



2Fo-Fc Map contoured at 0.6 σ

ID30B – Develop a flex plate gripper

Plate manipulator
nozzle



EMBL instrumentation team




Advanced instrumentation and software

MxCuBE



CRIMS



Crystal Direct™



ISPyB



ESRF

ID30B – To do

Immediate future

- Finalise Unipuck ‘double’ gripper integration in MxCuBE
- Finalise *in situ* plate screening
- Finalise MXPress automatic data collection WFs (as on MASSIF-1)

Near future

- Implement Flex plate gripper (prototype under construction)
- Develop plate functionalities in ISPyB/Exi ISPyB)
- Develop CRIMS-ISPyB connection for crystal ‘marking’
- Develop and implement dynamic beam size routines
- Add move to Si strip for low energy ranges

Future

- Improve long term beam stability
- Automate data collection from plates (MASSIF style)
- Expand lower energy ranges (<6 keV)
- Phase plate commissioning (D. de Sanctis)

Acknowledgements

EMBL Grenoble

Stephen Cusack
Regis Legrand

Matthew Bowler
Hassan Belrhali
Akim Khadrouche
Babu Manjasetty
Uli Zander

Florent Cipriani
Franck Felisaz
Robert Janocha
Marcos Lopez-Marrero
Gergely Papp
Christopher Rossi
Jeremy Sinoir
Clement Sorez

Josan Marquez
Guillaume Hoffmann
Zuzanna Kaczmarska

EMBL Hamburg

Thomas Schneider
Gleb Bourenkov
Ivars Karpics

ESRF

Jean Susini
Harald Reichert

Gordon Leonard
Sean McSweeney (NSLS II)
Christoph Müller-Dieckmann
Daniele de Sanctis
David Flot
Stephanie Monaco
Max Nanao
Didier Nurizzo
Sasha Popov
David von Stetten
Pascal Theveneau
Hugo Caserotto
Fabien Dobias
Thierry Giraud
Mario Lentini
John Surr
Antonia Beteva
Matias Guijarro
Olof Svensson
Alejandro de Maria
Solange Delageniere
Carole Clavel

UVHCI/IBS

Eva Pebay-Peyroula
Winfried Weissenhorn

Sandrine Vignon
Sylvie Costa
Corinne Iodice

